

SCIENTIFIC REPORTS

OPEN


Author Correction: Influence of environment on testing of hydraulic sealers

Mira Kebudi Benezra¹, Pierre Schembri Wismayer² & Josette Camilleri^{3,4}Correction to: *Scientific Reports* <https://doi.org/10.1038/s41598-017-17280-7>, published online 20 December 2017

In Table 3 of the Article the units $\mu\text{g}/\text{mm}^3$ are incorrectly given as mg/mm^3 and some negative values are incorrectly given as positive in the columns 'Solubility 28 Days' and 'Solubility'. In addition, the Solubility % value '<3' is incorrectly given as '>3'. The correct Table 3 appears below.

Material	Media	Test				
		Sorption 28 days	Solubility 28 days	Sorption 1 day	Solubility 1 day	Solubility
		$\mu\text{g}/\text{mm}^3$	$\mu\text{g}/\text{mm}^3$	$\mu\text{g}/\text{mm}^3$	$\mu\text{g}/\text{mm}^3$	%
AH Plus	water	1.7 ± 0.4	-0.3 ± 0.2	1.1 ± 0.1	0.6 ± 0.3	-0.04 ± 0.01
	HBSS	2.2 ± 1.0	0.4 ± 0.7	0.6 ± 0.1	0.4 ± 0.2	-2.3 ± 0.9
	DMEM	1.6 ± 0.6	-0.4 ± 0.2	1.2 ± 0.3	0.4 ± 0.1	-25.1 ± 10.8
BioRoot	water	32.6 ± 1.8	36.1 ± 4.4	31.1 ± 2.9	26.6 ± 2.0	15.8 ± 5.7
	HBSS	34.1 ± 0.6	39.4 ± 1.9	33.5 ± 1.8	29.1 ± 0.9	-30.2 ± 8.5
	DMEM	34.7 ± 3.2	41.4 ± 4.8	34.3 ± 0.6	31.1 ± 1.3	-31.1 ± 4.3
Endoseal	water	39.3 ± 2.7	37.5 ± 9.9	40.4 ± 1.5	17.9 ± 4.4	1.4 ± 0.3
	HBSS	39.0 ± 2.7	35.9 ± 3.1	41.4 ± 3.4	20.0 ± 2.7	-2.4 ± 1.7
	DMEM	44.1 ± 1.8	44.0 ± 6.6	38.5 ± 3.2	16.5 ± 5.2	-20.4 ± 7.3
ISO Standard		4049	4049	4049	6876	
value		<40	<7.5	<40	<7.5	<3

Table 3. Sorption and solubility values for test sealers using two standard methodologies (Mean \pm SD). *MTA Fillapex did not set so the fluid uptake could not be measured.

 **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2018

¹Department of Endodontics, Faculty of Dentistry, Bezmialem Vakif University, Istanbul, Turkey. ²Department of Anatomy, Faculty of Medicine and Surgery, University of Malta, Msida, Malta. ³Department of Restorative Dentistry, Faculty of Dental Surgery, University of Malta, Msida, Malta. ⁴School of Dentistry, University of Birmingham, Edgbaston, Birmingham, UK. Correspondence and requests for materials should be addressed to J.C. (email: J.Camilleri@bham.ac.uk)